



Certificate of Analysis

South China Special Gas Institute Co., Ltd
Foshan Huate Gas Co., Ltd

South China Special Gas Jinfeng Road, Heshun, Lishui Town, Nanhai, Foshan, Guangdong P.R.China, 528241 SOUTH CHINA T+86(0)757-85123400 F+86(0)757-85123403 CYLINDER NUMBER(S) *Indicates the actual cylinder(s) analyzed Production Date: 2010-09-17 Analysis Date: 2010-09-17 Expiration Date: 2011-09-17	To:		Certificate No: 20100917001		
	Product: Carbon Monoxide (CO)		Cylinder Size:		
	Grade: 2N5		Valve:		
			Contents:		
			Pressure:		
	Specification		Certified Concentration	Units	Analytical Method
	Carbon Monoxide (CO)	≥ 99.5	≥ 99.5	%	GC/PDID
	Oxygen (O ₂)	≤ 100	≤ 100	ppmv	GC/TCD
	Nitrogen (N ₂)	≤ 300	≤ 300	ppmv	GC/TCD
	Carbon Dioxide (CO ₂)	≤ 300	≤ 300	ppmv	GC/TCD
OHC	≤ 3500	≤ 3500	ppmv	GC/PDID	
Hydrogen (H ₂)	≤ 200	≤ 200	ppmv	GC/TCD	
Moisture (H ₂ O)	≤ 50	≤ 50	ppmv	Moisture Analyzer	
COMMENTS					

The provision of this certificate is expressly subject to and governed by Huate Gas Products China' Terms and Conditions of Sale available on our web site, www.scsgti.com

Analyst _____

Reviewer _____

The information and results contained herein should not be used for any purpose of advertisement and legal suitability